

We offer

Industrial temperature
sensors
Compensating cables
Glass thermometers
Temperature transmitters
Humidity transmitters
Pressure transmitters
Manometers
Valves
Indicators
Thermostats
Controllers
Relays
Recorders
Calibrators
Calibration furnaces
Heating elements

EU - DECLARATION OF CONFORMITY

Manufacturer: LIMATHERM SENSOR Sp. z o.o.
34-600 Limanowa, Skrudlak 1 Street

declares under the sole responsibility, that:

Head-type temperature sensors type:

..T..GB-..Exi-..., ..T..GN-..Exi-..., ..T..T-..Exi-..., ..T..P-..Exi-..., ..T..I-..Exi-...,
..T..SW-..Exi-..., ..T..SWG-..Exi-..., ..T..SWT-..Exi-..., ..T..GWN-..Exi-...,

meet the essential requirements of safety and health

Directive 2014/34/UE of 26 February 2014 and harmonized standards:

PN-EN 60079-0:2018, PN-EN 60079-11:2012, PN-EN 50303:2004,
PN-EN 60079-26:2015

what was confirmed by **EC-Type Examination Certificate KDB**
07ATEX055X dated 15.06.2007 with **Addendum no.1÷4** dated 4.11.2009,
16.06.2010, 3.03.2012 and 9.02.2015 and the Notification of a Quality Assurance
no. **GIG 18 ATEXQ070** dated 19.02.2018

issued by:

Central Mining Institute
Notified Body No. 1453
40-166 Katowice, Plac Gwarków 1

Above listed sensors obtain marking:

CE 1453 Ex I M1 Ex ia I Ma or
Ex II 2G Ex ia IIC T6-T1 Gb or
Ex II 1/2G Ex ia IIC T6-T1 Ga/Gb or
Ex II 1D Ex ia IIC T85 °C Da

Conditions of use:

Sensors installation and connection acc. to manual and safety
no. DTR-EExi-GL01.

Chairman of the Board
Chief Executive Officer


mgr inż. Zbigniew Juskiewicz

Limanowa, 12th March 2020



Certyfikat nr QS/67/20

District Court for
Cracow-City Center
XII Wydz. Gospodarczy
KRS 0000201394

Management
Zbigniew Juskiewicz
Capital
700 000 zł
VAT No: 737-19-66-189